

Statement of Compliance

Manufacturer's Name: _____ Dynisco Instruments _____

Product Name/Type: _____ Laboratory Melt Indexer (LMI) _____

Model Number(s): _____ LMI 5000 _____

The above product complies with the following:

- **EMC Directive (2004/108/EC)**

EN 61326-1:2006. Electrical equipment for measurement, control and laboratory use--EMC requirements. Part 1: General requirements EN 55011:2009+A1:2010, EN 61000-3.2:2006+A2:2009, EN61000-3-3:1995+A1:2001, EN 61000-4-2:1995+A1:1998 +A2:2001, EN 61000-4-3:2002, EN 61000-4-4:2004, EN 61000-4-5:1995+A1:2001, EN 61000-4-6:2009, EN 61000-4-11:2004

- **LV Directive (2006/95/EC)**

EN 61010-1:2010. Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1 General requirements.

(Support documentation may or may not accompany this statement)



3030 Gilchrist Road
Akron, Ohio, USA

Phone: (330) 745-1641
FAX: (330) 848-7326



CE Declaration of Conformity

Dynisco Instruments declares that the Model LMI 5000 Series Laboratory Melt Indexer complies with the essential requirements of the following EU Directives, based on the listed EU harmonized standards:

EMC Directive (2004/108/EC)

EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use--EMC requirements. Part 1: General requirements EN 55011:2009+A1:2010, EN 61000-3.2:2006+A2:2009, EN61000-3-3:1995+A1:2001, EN 61000-4-2:1995+A1:1998 +A2:2001, EN 61000-4-3:2002, EN 61000-4-4:2004, EN 61000-4-5:1995+A1:2001, EN 61000-4-6:2009, EN 61000-4-11:2004

LV Directive (2006/95/EC)

EN 61010-1:2010. Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1 General requirements.

I declare the above products meet the essential requirements that are specified above.

Patrick Kosuth
Vice President, Dynisco Engineering

6/10/2013

Date

Dynisco
P/N 975304
Rev A
05/31/2013